



Virginia Commonwealth University
VCU Scholars Compass

Theses and Dissertations

Graduate School

2004

An In Vitro Evaluation of the Elements Apex Locator Using the Endo Q System

Steven Andrew Brofsky
Virginia Commonwealth University

Follow this and additional works at: <http://scholarscompass.vcu.edu/etd>

 Part of the [Endodontics and Endodontology Commons](#)

© The Author

Downloaded from

<http://scholarscompass.vcu.edu/etd/1033>

This Thesis is brought to you for free and open access by the Graduate School at VCU Scholars Compass. It has been accepted for inclusion in Theses and Dissertations by an authorized administrator of VCU Scholars Compass. For more information, please contact libcompass@vcu.edu.

School of Dentistry
Virginia Commonwealth University

This is to certify that the thesis prepared by Steven A. Brofsky entitled AN IN VITRO
EVALUATION OF THE ELEMENTS APEX LOCATOR USING THE ENDO Q
SYSTEM has been approved by his or her committee as satisfactory completion of the
thesis or dissertation requirement for the degree of Master of Science.

B. Ellen Byrne, R. Ph, D.D.S., Ph.D. Virginia Commonwealth University School of Dentistry

James R. Lance, D.D.S. Virginia Commonwealth University School of Dentistry

Francis J. Robertello, DMD, MS, M.Ed. Virginia Commonwealth University School of Dentistry

B. Ellen Byrne, R. Ph., D.D.S., Ph.D. Interim Program Director Postgraduate Endodontics

Ronald J. Hunt, D.D.S., M.S. Dean of the School of Dentistry

Dr. F. Douglas Boudinot, Dean of the School of Graduate Studies

March 12, 2004

© Steven A. Brofsky 2004

All Rights Reserved

AN IN VITRO EVALUATION OF THE ELEMENTS APEX LOCATOR USING THE
ENDO Q SYSTEM

A Thesis submitted in partial fulfillment of the requirements for the degree of Master of
Science at Virginia Commonwealth University.

by

STEVEN A. BROFSKY
D.M.D, University of Connecticut, 1997
B.A. Psychology, S.U.N.Y at Stony Brook, 1993

Interim Director: B. ELLEN BYRNE, D.D.S., PHD
DEPARTMENT OF ENDODONTICS

Virginia Commonwealth University
Richmond, Virginia
March 2004

Acknowledgement

The authors would like to acknowledge Dr. Al Best for the statistical analysis and Drs. James Lance and Francis Robertello for their support and guidance. I would like to thank Dr. Gary R. Hartwell for his assistance in development of this project. Dr. Steven A. Brofsky would also like to thank Dr. Sean Fessenden for his help in preparation of this manuscript. This research project was funded by the Alexander Fellowship Fund

Dr. Steven A. Brofsky was a graduate endodontic resident at Virginia Commonwealth University School of Dentistry, Richmond, Virginia. He is currently in private practice limited to endodontics in Rockville, Maryland. Dr. B. Ellen Byrne is Chairman, Department of Endodontics, Virginia Commonwealth University School of Dentistry, Richmond, Virginia. Address requests for reprints to Dr. B. Ellen Byrne, Department of Endodontics, Box 980566, Virginia Commonwealth University School of Dentistry, Richmond, Virginia 23298-0566.

Table of Contents

	Page
Acknowledgements	ii
List of Figures	iv
Chapter	
1 Abstract	v
2 Introduction	1
3 Material and Methods	3
4 Results	5
5 Discussion	10
References	13
Appendices	
A	16
B	17
C	18

List of Tables

Page

Table 1: Comparison Between the Three Measurement Systems	7
---	---

List of Figures

Figure 1: Relationship Between Alginate Error and True Root Length.....8

Figure 2: Relationship Between Endo Q Error and True Root Length.....9

Abstract

AN IN VITRO EVALUATION OF THE ELEMENTS APEX LOCATOR USING THE ENDO Q SYSTEM

By Steven A. Brofsky, D.M.D.

A Thesis submitted in partial fulfillment of the requirements for the degree of Master of Science at Virginia Commonwealth University.

Virginia Commonwealth University, 2004

Interim Director: B. Ellen Byrne, D.D.S., PhD.
Chairman and Professor, Department of Endodontics

The objective of this study was two-fold: 1) to determine the accuracy of the Elements Apex Locator and 2) to compare the accuracy of the alginate and Endo Q models as in-vitro apex locator testing devices. Twenty teeth were decoronated at the CEJ and triplicate measurements were made using the Elements Apex Locator. All measurements were made to the apex reading of the apex locator. True length was established by visualizing the file tip at the apex with a dental operating microscope. Measurements were then taken with the teeth mounted in an alginate model and then in the Endo Q model. The two models were compared using a repeated-measure ANOVA. Statistically significant

differences occurred between the alginate and Endo Q models. The results showed that in 95% (n=19) of the cases, an accurate location to within ± 0.5 mm of the apical foramen was obtained with the Endo Q model and 55% (n=11) with the alginate model. With a ± 1.0 mm tolerance level, an accuracy of 95% (n=19) was found with the use of the alginate model. In conclusion, it seems that the Endo Q system was more suitable for testing the electronic apex locator than the alginate model. The Elements Apex Locator with the use of the Endo Q model was highly accurate in locating to within ± 0.5 mm of the apical foramen (Mean deviation = .17 mm).

Introduction

In 1942, Suzuki (1) determined the electrical resistance between an instrument in the root canal and an electrode applied to the oral mucous membrane to be constant. Twenty years later, Sunada (2) developed an electronic method that could measure the length of the canal according to this principle. First generation electronic apex locators (EAL) used direct current and the known constant resistance as a basis for working length determination. Second generation EALs used a single frequency alternating current (AC) and measured changes in impedance. Third generation EALs use multiple AC frequencies that monitors change in impedance. Recently, Sybron Endo has developed the Elements Apex Locator a fourth generation electronic apex locator. This unit uses multiple frequencies, but breaks impedance down into resistance and capacitance and measures those directly and independently during use.

There have been many in vitro models used to test the accuracy of electronic apex locators. Aurelio et al. (3) first suggested a model in 1983 that was found to accurately simulate the clinical setting. This model consisted of polystyrene culture tubes containing agar in phosphate-buffered saline. Donnelly (4) in 1993 substituted sugar-free Jell-O and saline for water, and plastic specimen cups. Others have used alginate as a method to test electronic apex locators (5-6).

An apex locator training device called the Endo Q (Acadental, Mission, KS) has been developed by Acadental. This device allows students to mount their extracted teeth

into a small box with materials developed to simulate the periodontal ligament of a tooth. To date there have been no studies published that test the accuracy of the Elements Apex Locator with the use of the Endo Q system. Therefore, the purpose of this study was two-fold: 1) to determine the accuracy of the Elements Apex Locator and 2) to compare the accuracy of the alginate and Endo Q models as in vitro apex locator testing devices.

Material and Methods

A group of twenty single rooted and single canal teeth with mature root apices and patent root canals were used. Extracted teeth were obtained and immediately placed into a solution of formalin. They were kept in this solution until needed. Each tooth was decoronated at the cemento-enamel junction to provide a flat horizontal surface for measurement purposes. The true length was determined by using a #15 stainless steel Flex-o-file (Dentsply/Mallefer, Tulsa, OK) placed into the root canal until the tip of the file reached the plane of the major diameter of the foramen as defined by Kuttler (7). A dental operating microscope (Global, St. Louis, MO) at the 1.0 step was used to view the file tip. The file length was determined by placing a silicone stopper flush to the flat horizontal coronal surface of the root when the file tip was placed to the level of the major foramen. The length of the file was then read using a millimeter digital caliper with a measurement accuracy of .01mm. Root canal measurements were made to the nearest .01mm. This measurement was repeated three times, the average of the three measurements was considered the true measurement to be used for statistical analysis and comparison purposes. Each tooth was then mounted in an alginate model as described by McDonald (8). The manufacturers recommended operating procedures for the Elements Apex Locator (Sybron Endo, Orange, Ca) were used. All measurements were to the “apex”

designation on the Elements unit or the apex location as visualized with the dental operating microscope.

After EAL measurements were recorded in triplicate, the teeth were removed from the alginate system with the use of a #557 fissure bur. After removal a ½ inch long piece of 3/8 inch tubing was placed over the apical portion of each root. The open portion of the tubing was then filled with the periodontal ligament simulating paste. The Endo Q-box was then placed into an oven preheated to 150° for 5 minutes. The Q-Box was then removed and the grounding plate or Q-link was inserted into the slot at the base of the Q-Box. A hole was then created with a mirror handle and the teeth were fit in the space. The fixing paste to be used to secure the teeth was then mixed and adapted around the teeth so that the level of material was flush with the top of the holder. The Endo Q with tooth was then allowed to cool for thirty minutes. Measurements were then taken in triplicate with the Elements Apex Locator to the “apex” designation.

Results

Twenty teeth were measured in triplicate using three measurement systems. The true root canal length was compared to the alginate and the Endo Q measurements. The three measurement systems were compared using a repeated-measure ANOVA. The results showed that the three measurement systems were significantly different ($F(2, 120) = 127$, $p\text{-value} < .0001$). The three group means were each significantly different from one another with the following ordering: Alginate > EndoQ > True. Note, however, that this result was not consistent from tooth to tooth ($F(38, 120) = 16$, $p > .0001$). In table 4 the means for each tooth and measurement system are shown. Next, comparisons between the measurement systems are shown separately for each of the 20 teeth. In the columns of p-values, the first value shows the overall test result for comparing the three measurements systems. If this p-value is significant (< 0.05), then the Bonferroni-adjusted p-value is shown for the three paired comparisons. The last column shows the means that are significantly different.

The most common result (six teeth out of 20) was that the three measurement systems were not significantly different. Next most common (five teeth out of 20) was the predominant ordering (Alginate>Endo Q>True). In three teeth (#4, #13, #15) the test indicated that

Alginate>Endo Q and Alginate>True but that Endo Q and True were not significantly different. There were other patterns as well.

Additionally, each of the two test-systems was compared to the true measurement to ascertain whether there was a trend between the amount of error and the true root length. In Figure 1, a reference line is shown at zero to indicate no measurement error and a trend line is shown to illustrate the relationship between the alginate error (the difference between the alginate measurement and the true measurement) and the true root length. As may be seen, the trend is not significant (p-value = .7). In figure 2, a comparable figure shows the relationship between Endo-Q error and true root length. Here, the relationship is only marginally significant (p-value = 0.0994).

The results showed that in 95% (n=19) of the cases, an accurate location to within ± 0.5 mm of the apical foramen was obtained with the Endo Q model and 55% (n=11) with the alginate model. With a ± 1.0 mm tolerance level, an accuracy of 95% (n=19) was found with both the Endo Q and the alginate model.

Tooth	Length			p-values				Order
	True	Alginate	Endo Q	difference	A = E?	A = T?	E = T?	
9	14.67	15.64	15.05	<.0001	<.0001	<.0001	0.0015	A>E>T
8	15.44	15.64	15.24	0.0011	0.0006	0.1638	0.2016	A>E
4	15.52	16.80	15.76	<.0001	<.0001	<.0001	0.0843	A>E, A>T
5	15.90	15.58	15.78	0.0099	0.1758	0.0081	0.7473	T>A
3	16.11	15.82	16.08	0.0140	0.0483	0.0243	1.0000	E>T>A
13	16.33	17.05	16.46	<.0001	<.0001	<.0001	0.7473	A>E, A>T
1	16.54	16.49	16.42	0.5482				
17	16.93	17.29	17.07	0.0043	0.1230	0.0033	0.6054	A>T
20	16.97	17.06	16.90	0.3377				
10	17.02	16.94	16.93	0.6683				
11	17.43	18.01	17.47	<.0001	<.0001	<.0001	<.0001	A>E>T
12	17.47	17.71	17.54	0.0840				
14	17.56	18.38	17.60	<.0001	<.0001	<.0001	<.0001	A>E>T
15	17.70	18.57	17.88	<.0001	<.0001	<.0001	0.2991	A>E, A>T
2	17.96	18.47	18.25	<.0001	0.1230	0.0000	0.0222	A>T, E>T
19	18.10	18.12	18.23	0.4046				
18	18.15	18.38	18.21	0.0921				
6	18.29	18.95	18.64	<.0001	0.0156	0.0000	0.0036	A>E>T
7	18.73	19.37	20.40	<.0001	<.0001	<.0001	<.0001	A>E>T
16	20.78	20.92	21.11	0.0106	0.2631	0.5427	0.0081	E>T>A
all	17.18	17.56	17.35	<.0001	<.0001	<.0001	<.0001	A>E>T

Table 1. Comparisons Between the Three Measurement Systems




FIG 1. Relationship between the Alginate Error and True Root Length




FIG 2. Relationship between the Endo Q Error and True Root Length

Discussion

Kuttler (7) and Green (9) have shown that the apex coincides with the anatomical foramen no more than 50% of the time. This limits the usefulness of radiographs, even if the quality is excellent (10). An excellent adjunct, therefore, is the use of an electronic apex locator. The $\pm 0.5\text{mm}$ to the apical foramen range has been considered as the strictest acceptable range (11,12). Thus, measurements attained within this tolerance are considered highly accurate. On the other hand, root canals do not always end with an apical constriction, a clear minor and major diameter or an apical foramen at the exact base of the cemental zone. This is why some authors prefer the $\pm 1.0\text{ mm}$ range as the acceptable range (13).

Studies indicate that the electrical resistance between the oral mucous membranes and the periodontium are equal. Studies also indicate that the resistance values between the two tissues are 6.5 ohms. According to the manufacturer the PDL paste with its pro-component and formula will create the resistance at 6.5 ohms between the PDL paste (PDL) and Q-link (lip). This could account for the improved accuracy of the endo Q over the alginate.

Although the alginate system is easy to use and relatively inexpensive compared to the Endo Q system, it quickly dries making it difficult to use long term. The Endo Q system takes more time to mount, but is still relatively easy to use. It does require a

heating device to heat up the Q-box, but provides much better stability of the tooth than the alginate. It can also be stored and used for a longer period of time without drying or cracking and provides realistic radiographic trabecular bone appearance.

The fourth generation Elements unit breaks impedance down into resistance and capacitance, and measures those directly and independently during use. This was done to help eliminate erroneous readings due to different combinations of these properties, which provide the same impedance reading. The Elements unit does not make any calculations internally as the third generation units do. Instead all combinations of capacitance and resistance relating to locations within the canal are calculated in advance and loaded into a matrix database within the unit. This decreases processing time as the unit only has to look up the information rather than going through a lengthy calculation. The result is that more data points can be averaged for each refresh of the display, which makes the displayed information much more stable. This is said to limit “jumpy or erratic readings on the display.

The Elements unit has an optional vitality scanner built in and a satellite display that can be clipped to the patient bib, instrument tray or microscope. At the time of data collection the unit could only be connected through the satellite system, since then an adapter now allows direct connection with the unit. The display has an apex designation and as the file is advanced down the canal a circle appears on the bottom left corner of the display. At .5mm the circular symbol will be complete and rotate.

In Conclusion, the results show the Endo Q system to be more suitable for testing the electronic apex locator than the alginate model. The Elements Apex Locator with the

use of the Endo Q model was highly accurate in locating to within $\pm 0.5\text{mm}$ of the apical foramen.

Literature Cited

Literature Cited

1. Suzuki K. Experimental study on iontophoresis. JPN J Stomatol 1942; 16:411-29
2. Sunada I. New Method for Measuring the Length of the Root Canal. J Dent Res 1962; 41:375-87
3. Aurelio JA, Nahmias Y, Gerstein H. A Model for Demonstrating an Electronic Canal Length Device. J Endodon 1983; 9:568-9
4. Donnelly JC. A Simplified Model to Demonstrate the Operation of Electronic Root Canal Measuring Devices. J Endodon 1993; 19:579-80
5. Kaufman AY, Keila S, Yoshpe M. Accuracy of a New Apex Locator: An In Vitro Study. Int End J 2002; 35:186
6. Katz A, Mass E, Kaufman AY. Electronic Apex Locator: A Useful Tool for Root Canal Treatment in the Primary Dentition. ASDC J Dent Child 1996; 1996; 63:414-7
7. Kuttler Y. Microscopic Investigation of Root Apexes. J Am Dent Assoc 1955; 50:544-52
8. McDonald NJ. The Electronic Determination of Working Length. Dent Clin N Am 1992;36:293-307.
9. Green D. Stereomicroscopic Study of 700 Root Apices of Maxillary and Mandibular Posterior Teeth. Oral Surg 1960; 13:728-33
10. Olson AK, Goerig AC, Cavataio RE, Luciano J. The Ability of the Radiograph to Determine the Location of the Apical Foramen. Int Endod J 1991; 24:28-35
11. Fouad AF, Krell KV, McKendry DJ, Koobusch GF, Olson RA. A Clinical Evaluation of Five Electronic Root Canal Length Measurement Instruments. J Endodon 1990; 16:446-9
12. Ricard O, Roux D, Bourdeau L, Woda A. A Clinical Evaluation of the Accuracy of the Evident RCM Mark II Apex Locator. 1991; 17:57-9

13. Gutmann JL, Leonard JE. Problem Solving in Endodontic Working Length Determinations. Continuing Education in Dentistry. 1995; 16:288-304

Appendix A

1	16.53	16.56	16.53
2	17.97	17.95	17.97
3	16.10	16.12	16.10
4	15.44	15.55	15.58
5	15.89	15.88	15.94
6	18.57	17.89	18.41
7	18.69	18.73	18.77
8	15.50	15.41	15.40
9	14.75	14.65	14.62
10	17.56	16.73	16.76
11	17.45	17.41	17.42
12	17.46	17.26	17.70
13	16.24	16.50	16.26
14	17.62	17.48	17.58
15	17.64	17.80	17.66
16	20.89	20.54	20.91
17	16.99	16.97	16.84
18	18.12	18.27	18.06
19	18.21	18.10	17.98
20	16.96	16.96	16.98

True Measurements in Triplicate.

Appendix B

1	19.45	19.54	19.48
2	18.45	18.45	18.52
3	15.80	15.86	15.80
4	16.16	17.12	17.11
5	15.58	15.57	15.58
6	19.12	18.86	18.86
7	19.34	19.42	19.35
8	15.64	15.64	15.65
9	15.63	15.66	15.63
10	16.98	16.98	16.87
11	18.00	18.02	18.01
12	17.76	17.74	17.62
13	17.05	17.05	17.05
14	18.35	18.35	18.44
15	18.57	18.54	18.61
16	20.90	20.94	20.93
17	17.27	17.27	17.33
18	18.35	18.37	18.41
19	18.12	18.10	18.13
20	18.37	18.31	18.34

Alginate Measurements in Triplicate

1	16.58	16.35	16.34
2	18.52	18.11	18.13
3	16.17	16.04	16.03
4	15.87	15.72	15.69
5	15.74	15.82	15.78
6	18.64	18.75	18.54
7	20.42	20.37	20.41
8	15.11	15.33	15.28
9	15.04	15.02	15.10
10	16.83	16.88	17.07
11	17.48	17.47	17.45
12	17.54	17.57	17.52
13	16.47	16.52	16.38
14	17.59	17.62	17.59
15	17.91	17.84	17.88
16	21.07	21.18	21.07
17	17.04	17.10	17.07
18	18.24	18.21	18.18
19	18.27	18.18	18.24
20	16.87	16.91	16.93

Endo Q Measurements in Triplicate

VITA

Steven A. Brofsky D.M.D.

Born in St. Vincents Hospital, Staten Island, New York November 12, 1970

Present Citizenship Richmond, Virginia, United States

Education:

July 2002 to July 2004:	Virginia Commonwealth University School of Dentistry, Richmond, VA Certificate in Endodontics Masters in Dentistry
August 1993 to May 1997:	University of Connecticut School of Dentistry, Farmington, CT D.M.D.
January 1989 to May 1993:	S.U.N.Y. at Stony Brook, Stony Brook, NY B.A. Psychology
September 1984 to May 1988:	Port Richmond High School Staten Island, NY College Preparation Emphasis

Honors:

2004 “An In Vitro Evaluation of the Elements Apex Locator System Using the Endo Q System” Publication pending, Presented at the American Association of Endodontics Annual Meeting.

1992, 1991, 1990 Deans List, S.U.N.Y. at Stony Brook, Stony Brook, NY.

Professional Affiliations:

American Association of Endodontists

American Dental Association